

## Declaration of Conformity

### Copal® G+C

#### Manufacturer

Heraeus Medical GmbH  
Philipp-Reis-Str. 8/13  
61273 Wehrheim, Germany

We herewith declare that all Requirements of Annex II of the Medical Device Directive 93/42/EEC are fulfilled.

Copal® G+C is a Class III product according to Annex IX, rule 13 of the Medical Device Directive.

Our Notified Body is TÜV Product Service, Ridlerstr. 65, 80339 München, Germany, CE 0123.

Wehrheim, 23.02.2009



Dr. André Kobelt  
Managing Director



i. V. Dr. Margarete Foth  
Safety Officer